

**Table 2G-7. Determining the Magnitude of Reference Scribe Line Offset**

<b>Probe/Angle-Measuring Device</b>	<b>Magnitude of <math>R_{SLO}</math></b>
Type S probe with inclinometer	$\theta_{null}$
Type S probe with protractor wheel and pointer	$90^\circ - \theta_{null}$
3-D probe with inclinometer	$90^\circ - \theta_{null}$
3-D probe with protractor wheel and pointer	$\theta_{null}$