

| | |
|----------------|-----------|
| SOURCE: | DATE: |
| MONITOR: | LOCATION: |
| SERIAL NUMBER: | SPAN: |

| | DAY | DATE | TIME | CALIBRATION VALUE | MONITOR RESPONSE | DIFFERENCE | PERCENT OF SPAN* |
|-----------------|-----|------|------|-------------------|------------------|------------|------------------|
| ZERO/ LOW LEVEL | 1 | | | | | | |
| | 2 | | | | | | |
| | 3 | | | | | | |
| | 4 | | | | | | |
| | 5 | | | | | | |
| | 6 | | | | | | |
| | 7 | | | | | | |
| HIGH LEVEL | 1 | | | | | | |
| | 2 | | | | | | |
| | 3 | | | | | | |
| | 4 | | | | | | |
| | 5 | | | | | | |
| | 6 | | | | | | |
| | 7 | | | | | | |

*Acceptance Criteria : $\leq 3\%$ of span each day for seven days.

Figure 2.2-1 Calibration Drift Determination